

QuickBlue



QuickBlue RealQuick PCR DNA Detection Kits

- QuickBlue RealQuick Salmonella
- QuickBlue RealQuick Listeria monocytogenes
- QuickBlue RealQuick Listeria innocua
- QuickBlue RealQuick Staphylococcus aureus
- QuickBlue RealQuick Enterobacter sakazakii
- QuickBlue RealQuick Escherichia coli
- QuickBlue RealQuick Campylobacter jejuni
- QuickBlue RealQuick Clostridium perfringens
- QuickBlue RealQuick Vibrio vulnificus
- QuickBlue RealQuick Vibrio parahaemolyticus
- QuickBlue RealQuick Vibrio alginolyticus
- QuickBlue RealQuick Vibrio cholerae
- QuickBlue RealQuick Sus scrofa
- QuickBlue RealQuick Legionella pneumophila

Q-Bioanalytic GmbH
Fischkai 1
D-27572 Bremerhaven, Germany
Tel. +49 (0) 471 - 48 32 440

QuickBlue



Listeria monocytogenes

Real-Time PCR DNA Detection Kit

User manual

Index	Page	Substances	Order No: QB-RT-01-50
Substances	2	QuickBlue DNA Detection Kit Components for 25 µl reaction	50
Additional required material	2		
Storage	2		
Overview PCR detection	3	1. PCR premaster (white) contains:	1 Cup
Specifications	4	•RNase free water •PCR buffer •MgCl ₂ (1,5 mM) •dNTP-Mix including dUTP •primer pair •HotMaster Taq polymerase	
Product limitations	4		
Introduction	5		
Protocol	5		
Mastermix preparation	6		
Prearrangement	6	2. Uracilglycosylase (UNG) (yellow)	1 Cup
PCR program	6	3. TaqMan probe (green)	1 Cup
Pipetting of PCR mastermix and DNA samples	7	4. Reference DNA (positive control) (red)	1 Cup
Thermocycling	7	5. RNase free water (blue)	1 Cup
PCR results	9	6. User manual	1
Trouble shooting	10		

Additional required material:

Instrumentation and media for preculture (if necessary), equipment for water-filtration (if necessary), PBS buffer, DNA extraction kits, Instruments for PCR (Realtime cycler).

Storage:

The QuickBlue DNA Detection PCR Kit should be stored at - 20 °C , short storage at + 5 °C is possible. Correct storage and handling increases the stability of the PCR kit!

Terms an Conditions

- a) The purchase of the goods only conveys to Buyer the licence to use the purchased goods according to the product manual and subject to the following licence conditions.
- b) Q-Bioanalytic's goods are designed exclusively for use in scientific research and development. Any use of Q-Bioanalytic's goods for in vitro diagnostic purposes, or as pharmaceuticals shall not be permitted.
- c) **Any copying, disassembling, analyzing, sequencing or other form of reverse engineering of all Q-Bioanalytic's goods is expressly forbidden.**
- d) Any resale of the goods delivered by Q-Bioanalytic is only permitted if the Buyer agrees with its customer to bind the customer to terms identical to this item a) – e) in favor of Q-Bioanalytics.
- e) Q-Bioanalytics hereby assigns any rights, including, but not limited to patents and patent applications, that it may obtain due to a violation of it obligations under item a), b), c) or d) to Q-Bioanalytic, which accepts such assignment. Furthermore, the buyer shall pay Q-bioanalytic a royalty in the amount of 30% of all the net sales of products developed due to such violation. Furthermore, the buyer shall indemnify and hold harmless the seller of any third parties claims based on such violation, and shall be liable for any damages of Q-Bioanalytic insofar as they have not been cured by the aforesaid provisions.

Trouble Shooting

Problem	Comments and suggestions
No fluorescence with the positive control	Pipetting failure, positive DNA degraded, probe missing
PCR product in negative control	UNG missing, pipetting failure with genomic sample DNA, PCR premaster or mastermix contaminated during preparation
To little amount of components in the PCR Kit	Pipetting failure in previous PCR preparations, so the volumes of the components will not suffice
No curves	Wrong filter system: 6-Fam filter systems were not chosen

Overview PCR detection:

If applicable:
preculture (A) /
water-filtration (B)

Bacteria DNA
preparation

(A) Pre-enrichment of sample in buffered peptone water or appropriate selective medium (1:10 w/v: 25 g sample + 225 ml sterile peptone-water) over-night at convenient temperature.
(B) Filtration of 100 ml with an appropriate filter system* (0,45 µm!), e.g. cellulose nitrate filter.

(A) Use QuickBlue DNA purification kit or other appropriate DNA purification methods. Thermal lysis may only be used in matrices which does not contain components inhibiting the PCR reaction.

QuickBlue-Kit

DNA purification

bacteria: (A) after DNA preparation dilute DNA 1:10 in TE buffer to avoid PCR inhibition.

(B) extract DNA from the filter directly (e.g. with CTAB lysis buffer*);
eucaryota: use a commercial DNA purification kit* like column systems or CTAB extraction.

Mastermix
preparation

combine premaster (white), TaqMan probe (green) and UNG (yellow), amounts see table 3 (page 9), mix well (= Mastermix).

PCR
pipetting

Pipetting: 24 µl Mastermix
and 1 µl DNA
or 1 µl H₂O (= negative control, blue)
or 1 µl reference DNA (= positive control, red)

Thermocycling

Initialphase: 2 min 50 °C, 10 min 95 °C,
40 cycles: 20 s 95 °C,
30 s 60 °C,
20 s 72 °C
End phase: hold 10 °C

Specifications

Specificity

Listeria monocytogenes

The primer and probe systems are species specific for the respective **QuickBlue** PCR DNA Detection Kits.

Sensitivity

After pre-enrichment < 1 cells / 25 g

Without pre-enrichment < 10 cells / g or 10 non-degraded genomes

Technical support


The Q-Bioanalytic GmbH offers a qualified technical support. If you have questions or problems with the handling of the PCR detection kit, do not hesitate to contact us. For our technical support and more information about our QuickBlue PCR Detection Kit please call phone.: +49 - (0)471 - 48 32 440.

Product limitations

QuickBlue DNA Detection Kit is licensed for food testing only.

Safety information

When working with the QuickBlue DNA PCR Detection Kit, protective clothing should be worn as well as lab gloves. Work should be done under PCR-sterile conditions. Avoid contact with DNA or DNA contaminated areas and surfaces during PCR mastermix preparation. First pipette the mastermix and subsequently the DNA into the PCR tubes or PCR plates.



curves
no 1 – 7:
serial dilution

negative
Controls with
non target
species

Figure 2: Demonstration of RT-PCR results performed with a LightCycler® 480 instrument. A serial dilution with reference DNA (curve 1 – 7) and the negative controls is shown.

The curves no 1 - 7 show a dilution series of *Listeria monocytogenes* DNA. The bottom lines are negative controls and DNA of non target species.

The first amplification signal can appear with the 14th cycle. This depends on the purity of DNA (→ inhibition by matrix components). Signals after cycle 38 should be repeated with higher concentrated sample.

Table 3: Preparation of premaster for parallel approaches:

Number:	H ₂ O:	Premaster:	Probe:	UNG:	Number:	H ₂ O:	Premaster:	Probe:	UNG:
1	11	12.5	1	0.2	49	539	612.5	49	9.8
2	22	25	2	0.4	50	550	625	50	10
3	33	37.5	3	0.6	51	561	637.5	51	10.2
4	44	50	4	0.8	52	572	650	52	10.4
5	55	62.5	5	1	53	583	662.5	53	10.6
6	66	75	6	1.2	54	594	675	54	10.8
7	77	87.5	7	1.4	55	605	687.5	55	11
8	88	100	8	1.6	56	616	700	56	11.2
9	99	112.5	9	1.8	57	627	712.5	57	11.4
10	110	125	10	2	58	638	725	58	11.6
11	121	137.5	11	2.2	59	649	737.5	59	11.8
12	132	150	12	2.4	60	660	750	60	12
13	143	162.5	13	2.6	61	671	762.5	61	12.2
14	154	175	14	2.8	62	682	775	62	12.4
15	165	187.5	15	3	63	693	787.5	63	12.6
16	176	200	16	3.2	64	704	800	64	12.8
17	187	212.5	17	3.4	65	715	812.5	65	13
18	198	225	18	3.6	66	726	825	66	13.2
19	209	237.5	19	3.8	67	737	837.5	67	13.4
20	220	250	20	4	68	748	850	68	13.6
21	231	262.5	21	4.2	69	759	862.5	69	13.8
22	242	275	22	4.4	70	770	875	70	14
23	253	287.5	23	4.6	71	781	887.5	71	14.2
24	264	300	24	4.8	72	792	900	72	14.4
25	275	312.5	25	5	73	803	912.5	73	14.6
26	286	325	26	5.2	74	814	925	74	14.8
27	297	337.5	27	5.4	75	825	937.5	75	15
28	308	350	28	5.6	76	836	950	76	15.2
29	319	362.5	29	5.8	77	847	962.5	77	15.4
30	330	375	30	6	78	858	975	78	15.6
31	341	387.5	31	6.2	79	869	987.5	79	15.8
32	352	400	32	6.4	80	880	1000	80	16
33	363	412.5	33	6.6	81	891	1012.5	81	16.2
34	374	425	34	6.8	82	902	1025	82	16.4
35	385	437.5	35	7	83	913	1037.5	83	16.6
36	396	450	36	7.2	84	924	1050	84	16.8
37	407	462.5	37	7.4	85	935	1062.5	85	17
38	418	475	38	7.6	86	946	1075	86	17.2
39	429	487.5	39	7.8	87	957	1087.5	87	17.4
40	440	500	40	8	88	968	1100	88	17.6
41	451	512.5	41	8.2	89	979	1112.5	89	17.8
42	462	525	42	8.4	90	990	1125	90	18
43	473	537.5	43	8.6	91	1001	1137.5	91	18.2
44	484	550	44	8.8	92	1012	1150	92	18.4
45	495	562.5	45	9	93	1023	1162.5	93	18.6
46	506	575	46	9.2	94	1034	1175	94	18.8
47	517	587.5	47	9.4	95	1045	1187.5	95	19
48	528	600	48	9.6	96	1056	1200	96	19.2

Introduction

Target organisms will be detected with species specific oligonucleotides (specific primer from German LFGB § 64 or NCBI database entries). High specificity and sensitivity of the PCR kit result from chosen PCR components as well as special Taq-Enzyme and PCR buffer. The QuickBlue DNA Detection PCR Kit was optimized on several thermocyclers. The user manual contains a general protocol for this PCR application. Kits were validated with Roche LightCycler 480, Applied Biosystems StepOne, Stratagene MX 4000 and Eppendorf RealPlex Cyclers.

Protocol

After appropriate pre-enrichment and/or DNA preparation perform the following steps for the PCR:

- Preparation of mastermix
- Pipetting the samples and controls
- Thermocycling
- Evaluation

Important advices for mastermix preparation

- Start the PCR only with the PCR program as shown in table 1 (page 6)
- Components are optimized for a volume of 25 µl
- Pipette always 1 µl of your DNA sample to the specific mastermix to a total volume of 25 µl
- Use only declared PCR tubes and PCR plates (from polyethylen, polypropylene or polycarbonate, no polystyrene!!!) for this PCR kit.

Mastermix preparation

Prearrangement

The PCR components have to thaw in a sterile, DNA-free and PCR product free area (→ separate PCR room or PCR hood). Place the pipettes, tips and waste container closely to your workbench so you could not touch other objects with your disposable lab gloves during pipetting. This is to avoid contaminations of your PCR components. Set the PCR tubes/plate and the DNA samples in a separate laboratory or aside the PCR hood.

PCR program

Before pipetting make sure that following PCR program is on your thermocycler:

Table 1: Demonstration of the specific PCR program

PCR steps	Time	Temperature	Commentary
UNG activity	2 min	50 °C	eliminates contaminations
Initial denaturation	10 min	95 °C	activates HotStart enzyme
Number of cycles		x 40	defines product amount
Denaturation	20 s	95 °C	denaturation of dsDNA
Primer annealing	30 s	60 °C	primer annealing on ssDNA, fluorescence measurement (6-Fam filter system, 483 nm)
Elongation	20 s	72 °C	primer extension to PCR product
storage	hold	10 °C	storage temperature

Pipetting the PCR mastermix and DNA samples

Table 2: Demonstration of the composition (per PCR reaction)

Components	Volume / reaction	End concentration
RT-PCR premaster (white)	12,5 µl	-
RNase free water (blue)	11,0 µl	-
UNG (yellow)	0,2 µl	0,2 Unit
Probe (green)	1,0 µl	-
Mix volume	24,7 µl	
DNA	1,0 µl	variable
Total volume	25,7 µl	-

1. Depending on your total PCR amount pipette the adequate volumes of premaster, water, Uracil-Glycosylase (UNG) and probe in a sterile PCR tube (e.g.: 1,5-ml reaction cup).
2. Mix the mastermix on a shaker, to disperse components homogeneously.
3. Put the PCR components back into the freezer (at - 20 °C).
4. Aliquot 24 µl of your mastermix into the labeled PCR tubes (0,2 ml tubes, 8-module-strips, or 96/384-PCR-plate). Don't forget controls!
5. Pipette 1 µl of your sample DNA per cavity into the mastermix. This avoids crosscontaminations!
6. Don't forget to pipette 1 µl of the negative control (blue cup) and 1 µl of the positive control (red cup) into separate PCR mastermixes. Include a positive control reaction in your DNA preparation to exclude effects of inhibition by the matrix analysed.
7. Close the tubes or plates well and put them into the thermocycler.

Thermocycling

Start the specific PCR program (see table 1, page 6).